NM3 / NM2B O.R. DATA MANAGER SOFTWARE INSTALLATION PROCEDURE

INSTALLATION PROCEDURE

CAUTION: To avoid losing data, never reset the O.R. DATA MANAGER when the disk drive indicator light is on.

NOTE: The following precautions must be observed to ensure that the disk is not damaged or the data corrupted during use or storage:

- Keep disks away from dust and dirt. Small particles of dust or dirt may scratch the magnetic surface and destroy data.
- Store disks in a moderate environment such as normal room temperature and humidity.
- Never touch the magnetic surface of the disk.
- Store disks in a diskette container when they are not in use.
- Never wipe, brush or try to clean the magnetic surface in any way.
- Label all disks for proper identification.
- Do not use an eraser on a label affixed to a disk.
- Remove the disk before unplugging the O.R. DATA MANAGER.

1. Ensure that the O.R. DATA MANAGER is plugged into a live AC receptacle.
2. Press the eject button on the disk drive to ensure that the disk drive is empty. See Figure 1.
3. Hold the disk so that the metal media cover is facing toward the disk slot and the round metal hub is facing toward the eject button and indicator light.
4. Insert the disk into the disk drive, pressing it in until it locks into place.
5. Simultaneously press the CONTROL, ALT and DELETE ENTRY keys on the keyboard.
6. Observe the remote display during the procedure. Downloading is complete when the message "REMOVE THE FLOPPY DISK FROM THE FLOPPY DRIVE AND DEPRESS ANY KEY" appears on the remote display.
7. Press the eject button on the disk drive. The disk will partially self-eject from the drive.
8. Remove the disk from the disk drive.

INSTALLATION PROCEDURE (continued)

9. Verify that the software has been properly downloaded: On the keyboard, press and hold the ALT key, and press the V key. The current software version number shown on the remote display should agree with that marked on the program disk or its accompanying documentation.



Figure 1: O.R. DATA MANAGER FRONT VIEW



Figure 2: O.R. DATA MANAGER REAR VIEW

NORTH
AMERICAN
DRÄGER

Quality Service for Life®

Technical Service Department
3122 Commerce Drive
Telford, PA 18969
(215) 721-5402
(800) 543-5047
(215) 723-5935 Fax

Part Number: SP00082
Rev: B
Date: May 18, 1994
© N.A.D. Inc. 1994